



Bio-X Diagnostics

ANTI-KHV MONOCLONAL ANTIBODY (2)

BIO 354

(Reagent for indirect immunofluorescence or immunoperoxidase assay)

**REAGENT FOR DETECTING KOI HERPES VIRUS
ON CELL CULTURES.**

I – INDIRECT IMMUNOFLUORESCENCE ASSAY PROCEDURE

Fix the cell culture for 10 minutes at -20°C using the following fixative:

- Acetone 1 volume + Ethanol 1 volume

Then rinse with PBS.

Dilute the reagent twentyfold in PBS prepared according to the following formula:

Phosphate Buffer Saline (PBS)	
NaCl:	8 gm
KH_2PO_4 :	0.2 gm
KCl:	0.2 gm
$\text{Na}_2\text{HPO}_4 \cdot 2\text{H}_2\text{O}$:	1.15 gm
NaN_3 :	0.1 gm
H_2O	1 L

Incubate the preparation on the sample for 1 hour at $21^{\circ}\text{C} \pm 3^{\circ}\text{C}$, preferably in a humidity chamber.

Upon completion of this incubation period rinse the preparation with a PBS solution.

Then add the conjugate (fluorescein-labelled anti-mouse immunoglobulin) at the manufacturer's recommended dilution. The conjugate available from Bio-X Diagnostics (BIO 156) should be diluted twentyfold in PBS-Evans Blue solution.

PBS-Evan Blue	
PBS	1 liter
Evans Blue:	0.01 gm

Incubate the preparation on the sample for 1 hour at 21°C +/- 3°C , preferably, in a humidity chamber.
After this second incubation step rinse the preparation with PBS.

Dry the slide, then add the mounting medium made up as follows:

Mounting medium	
Glycerol	9 parts by volume
PBS	1 part by volume

Place a cover slip on the slide, then observe under a microscope fitted for fluorescence detection.

The antibody may be kept in its original vial at 4°C for more than a year. Never freeze this reagent.
Once diluted in PBS, the antibody remains stable for one week between +2°C and +8°C.

II – INDIRECT IMMUNOPEROXIDASE ASSAY PROCEDURE

Fix the cell culture for 15 minutes at 21°C +/- 3°C using one of the following fixatives:

- 9:1 (v/v) acetone/water solution

Then rinse with PBS.

Dilute the reagent twentyfold in PBS prepared according to the following formula:

PBS	
NaCl:	8 gm
KH ₂ PO ₄ :	0.2 gm
KCl:	0.2 gm
Na ₂ HPO ₄ . 2H ₂ O:	1.15 gm
NaN ₃ :	0.1 gm
H ₂ O	1 L

Incubate the preparation on the sample for 1 hour at 21°C +/- 3°C , preferably in a humidity chamber.

Upon completion of this incubation period rinse the preparation with PBS.

Then add the conjugate (peroxidase-coupled anti-mouse immunoglobulin) at the manufacturer's recommended dilution. The conjugate available from Bio-X Diagnostics (BIO 157) should be diluted twentyfold in PBS.

Incubate the preparation on the sample for 1 hour at 21°C +/- 3°C , preferably in a humidity chamber.

After this second incubation step rinse the preparation with PBS.

Then add the chromogen (AEC, precipitating TMB, DAB, etc.) and the substrate (hydrogen peroxide) according to the manufacturer's instructions. Examine under the microscope for the presence of the coloured marker.