



Instruction manual ADC24EPC_NEOC_NO_(EN)_V02 08/2024

Extraction Positive Control NEOSPORA

Reference: ADC24EPC

Neospora caninum positive control for extraction control

For veterinary in vitro use only



Kit composition

Content		ADC24EPC Kit 100 reactions
EPC NEO	Extraction Positive Control Neospora caninum	1 lyophilized vial (To reconstitute)
NF-Water	Nuclease free water	1 x 500 μL tube (Ready to use)

Associated PCR kits

Associated kits	References
ADIAVET™ NEOSPORA REAL TIME	418023

Revision history

Date	Version	Modifications
06/2023	V01	First version
08/2024	V02	Modification of EPC NEO calibration (1 to 10 x LD METHOD)

Note: minor typographical, grammar and formatting changes are not included in the revision history.

A. Test principle

The extraction positive control is made with an inactivated culture of E.coli containing the target of Neospora caninum.

This extraction positive control after rehydration can be used for two purposes:

- the extraction positive control can be used within each extraction series and allows the day-to-day extraction step
- it can be used to confirm the LOD_{METHOD}. It is calibrated between 1 and 10 the LODMETHOD.

It is the responsibility of the laboratory to obtain the matrix free of Neospora caninum.

B. Storage conditions

On receipt, the kit should be stored dry at +2/8 °C.

Reconstituted reagents should be stored at <-15 °C until the expiration date.

Store away from sunlight.

Do not mix reagents of two different batches.

C. Material required but not provided

- Real-time Thermal cycler and device.
- Instrument for homogenous mixing of tubes.
- Pipettes of 1 10 μ L, 20 200 μ L and 200 1000 μ L.
- Nuclease-free filtered pipette tips.
- Nuclease-free microtubes of 1,5 mL and 2 mL.
- Powdered-free latex or nitrile gloves.
- Nuclease-free water.
- Associated PCR kits.
- Kit for nucleic acids extraction.
- Matrix free of Neospora caninum.

D. Warnings and precautions

- For veterinary in vitro use only.
- For animal use only.
- For professional use only.
- All instructions must be read before performing the test and strictly respected.
- Do not use reagents after the expiration date.
- Do not use reagents if the packaging is damaged.
- Do not open PCR wells or tubes after amplification.
- Do not mix reagents from different batches.
- Used material must be disposed of in compliance with the legislation in force regarding environmental protection and biological waste management.
- This kit contains products of animal origin. Certified knowledge of the origin and/or sanitary state of the animals does not totally guarantee the absence of transmissible pathogenic agents. It is therefore recommended that these products be treated as potentially infectious and handled observing the usual safety precautions (do not ingest or inhale).

E. Procedure

1. Reconstitution of the control

- Add 500 uL of « NF Water » in « EPC NEO » vial.
- Vortex vigorously at least 20 seconds.
- After reconstitution, the control is calibrated between 1 et 10 X LD_{METHOD}.
- Aliquot and store at <-15°C. Do not defrost more than 3 times.

2. LODMETHOD confirmation

a. Extraction

Add **5 µL** of « **EPC NEO** » to the matrix free of *Neospora caninum*, then make the extraction on duplicate.

Two independent sessions should be realized.

b. Amplification

Extracted nucleic acids are amplified with the associated Bio-X Diagnostics PCR kits according to kit instructions.

c. Validation of results

The four results obtained with « EPC NEO » must be positive.

3. Extraction step control

a. Extraction

Add 5 µL of « EPC NEO » to the matrix free of Neospora caninum. then make the extraction.

b. Amplification

Extracted nucleic acids are amplified with the associated Bio-X Diagnostics PCR kits according to kit instructions.

c. Validation of results

All the obtained results during the different series of extraction constitute the control chart.

Bibliography

U47-600-1: Méthodes d'analyse en santé animale - PCR (réaction de polymérisation en chaîne) - Partie 1: exigences et recommandations pour la mise en oeuvre de la PCR en santé animale.

Symbols

Symbol	Signification
REF	Catalog number
	Manufacturer
¥	Temperature limitation
Σ	Use by
LOT	Batch code
[]i	Consult Instructions for Use
Σ	Contain sufficient for "n" tests
VET	For veterinary in vitro use only – For animal use only
类	Keep away from sunlight



Contact us