

Protocols for KingFisher instruments
Using “ADIAMAG”
extraction kit

Ref. NADI003 (200 extractions)

Ref. NADI003-XL (800 extractions)

Using “ADIAMAG”

extraction kit

MAIN CHANGE SINCE PREVIOUS VERSION	3
I. GENERAL INFORMATION	4
1. Purpose of the kit.....	4
2. Description of test.....	5
II. MATERIAL & REAGENTS	7
1. Composition of kit	7
2. Validity and storage.....	7
3. Equipment required, but not supplied in the kit.....	7
III. USE OF SAMPLES AND CONTROLS	9
1. Precautions.....	9
2. Storage of nucleic acid extracts.....	9
3. Controls preparation	9
IV. PREPARATION OF THE BUFFERS AND LOADING OF THE INSTRUMENTS.....	10
1. Preparation of the binding buffer	10
2. Loading of the instrument.....	10
3. Instrument programs	11
V. SAMPLES PREPARATION BEFORE TRANSFER ON THE DEVICE.....	12
1. From blood/sera	12
2. From oral fluids	12
3. From milk	13
4. From culture supernatant.....	13
5. From ear notch samples and humid TST-L from ALLFLEX ear notch (BAH).....	13
6. From conservation buffer of TST-L from ALLFLEX ear notch (BAL).....	14
7. From skin biopsy.....	14
8. From trachea- bronchial washing	14
9. From water and urine.....	15
10. From coelomic liquid.....	15
11. From foetal gastric fluid.....	15
12. From feather	16
13. From FTA Card.....	16
14. From tissue or brain.....	17
15. From swab	19
16. From faeces	21
17. From environmental sample (drag swab).....	23
18. From sperm or egg fish.....	24
VI. AMPLIFICATION.....	25
VII. INDEX OF SYMBOLS	26

Main change since previous version

N/A	Not Applicable (first publication)
Correction	Correction of document anomalies
Technical change	Addition, revision and/or removal of information related to the product
Administrative	Implementation of non-technical changes noticeable to the user
Note: minor typographical, grammar and formatting changes are not included in the revision historic	

Release Date	Part Number	Change type	Change summary
2021/04	NEKF-22	Technical change	Addition of protocols for PRV from blood, swab and tissues. Addition of protocols for <i>Besnoitia besnoiti</i> from skin biopsy.
2022/01	NEKF-23	Administrative	Appendix 1 Removal
2022/01	NFKF-23	Technical change	Protocol adjustment for the search of <i>Besnoitia besnoiti</i> from skin biopsy. Addition of protocol for detection of <i>Coxiella</i> , and <i>Chlamydia</i> from foetal gastric fluid. Addition of protocols for Marek virus. Addition of protocols for detection of avian influenza from feather and FTA card. Addition of protocol for detection of avian mycoplasma from FTA card. Addition of protocols for detection of <i>Salmonella</i> from abortion samples and foetal gastric fluid. Addition of protocols for PCV2 and PCV3.
2022/09	NEKF-24	Technical change	Addition of protocols for the search of BVDv from humid TST-L from ALLFLEX ear notch (BAH) or conservation buffer of TST-L from ALLFLEX ear notch (BAL)
2022/11	NEKF-25	Technical change	Addition of protocol for the search of <i>Mycobacterium avium</i> subsp <i>paratuberculosis</i> from faeces, tissue. Addition of protocol for search of Influenza virus and PRRS virus from oral fluids. Addition of environmental samples matrix for the search of Influenza Virus. Addition of a protocol for search of Bronchitis Virus.
2023/11	NEKF-26	Technical change	Addition of protocols for the search of IHNv and VHSV from sperm/egg fish. Addition of protocols for the search of EHDV from blood and tissue. Addition of protocols for the search of BTv from tissue.

I. General information

1. Purpose of the kit

ADIAMAG is a nucleic acids DNA and RNA extraction kit based on nucleic acids adsorption on magnetic beads. The kit is adapted for KingFisher™ mL, DUO and 96/Flex instruments and contains all the buffers required to perform DNA/RNA extraction from different matrices.

If using a different automaton than those validated, please contact support.pcr@biox.com.

After samples lysis (performed separately from the instrument and using appropriate buffers), the supernatants can be handled with the same automated purification process using KingFisher™ mL, KingFisher DUO or KingFisher™ 96/Flex instruments, irrespective of the pathogen and matrix used. The following table describes the protocol used with the different instruments.

Well/plate	Description
1	Nucleic acids release and binding to magnetic beads. Beads collection by magnetic rods and transfer into the step 2.
2	Beads release. W3 buffer washing step. Beads collection by magnetic rods and transfer into the step 3.
3	Beads release. W4 buffer washing step. Beads collection by magnetic rods and transfer into the step 4.
4	Beads release. Ethanol 80 % washing step. Beads collection by magnetic rods, air drying step followed by transfer into the step 5.
5	Nucleic acids elution step with E6 buffer. Nucleic acid-free beads collection and transfer into the step 3.

Contact us to obtain the files for your KingFisher™ instruments.

2. Description of test

Bio-X Diagnostics validated the ADIAMAG kits with various matrices for the detection of different DNA/RNA pathogens in combination with the ADIAVET™ et ADIALYO™ range. The following table summarises the validated protocols.

		Samples																	
		EDTA blood	Sera, plasma	Ear notch	Skin biopsy	Tissue	Brain	Swab	Oral fluid	Faeces	Milk	tracheo –bronchial washing	Culture supernatant	Water/urine	Semen	Feather	FTA card	Foetal gastric fluid	Environmental sample
Animal diseases	Influenza virus					X		X	X	X		X	X			X	X		X
	Chlamydia					X		X			X							X	
Ruminant diseases	BoHV-4							X											
	Coxiella burnetii					X		X		X	X							X	
	Leptospira					X		X					X						
	Neospora caninum					X	X	X											
	BVD virus	X	X	X		X*					X								
	Schmallenberg virus (SBV)	X	X			X	X												
	Anaplasma phagocytophilum	X				X		X											
	Besnoitia besnoiti	X			X														
	Toxoplasma gondii					X	X	X		X									
	Bluetongue virus (BTV)	X				X													
	Epizootic Haemorrhagic disease virus (EHDV)	X				X													
	M. avium subsp. paratuberculosis					X				X									X
	Salmonella					X		X										X	
Swine diseases	Mycoplasma hyopneumoniae					X		X	X			X							
	Virus PRV, Aujeszky's disease					X*	X	X				X							
	PRRS virus	X	X			X		X	X			X							
	Actinobacillus pleuropneumoniae					X		X											
	Classical Swine Fever Virus (CSFV)	X	X			X*													
	African Swine Fever Virus (ASFV)	X	X			X*		X											
	Porcine Circovirus (PCV2&PCV3)	X	X			X													

*If using the reference NADI003-XL, it is necessary to add the reference NADI004.

For the pool size, refer to the user manual of the kit concerned.

		Samples																	
		EDTA blood	Sera, plasma	Ear notch	Skin biopsy	Tissue	Brain	Swab	Oral fluid	Faeces	Milk	tracheo –bronchial washing	Culture supernatant	Water/urine	Sperm, egg, coelomic liquid	Feather	FTA card	Foetal gastric fluid	Environmental sample
Avian diseases	Mycoplasma gallisepticum M. synoviae M. meleagridis M. iowea							X									X		
	Ornithobacterium rhinotracheale							X											
	Marek virus					X			X							X			X
	Bronchitis Virus							X											X
Contagious equine metritis (CEM)	Taylorella equigenitalis, Taylorella asinigenitalis, Klebsiella pneumoniae Pseudomonas aeruginosa							X											
Fish diseases	Infectious hematopoietic necrosis virus (IHNV)					X						X		X					
	Viral hemorrhagic septicemia virus (VHSV)					X						X		X					
	Infectious pancreatis necrosis virus (IPNV)					X						X		X					
Feline disease	Feline coronavirus (FIP)	X						X											

II. Material & reagents

1. Composition of kit

The ADIAMAG kits contain the following buffers:

Designation of reagents	Type of reagent	NADI003 ADIAMAG 200 extractions	NADI003-XL ADIAMAG XL 800 extractions	NADI004 LB3 -125mL	Instruction
Lysis Buffer - LB1	Lysis buffer 1	1 x 20 mL	1 x 100 mL		Ready to use
Lysis Buffer - LB2	Lysis buffer 2	1 x 50 mL	2 x 125 mL		Ready to use
Lysis Buffer - LB3	Lysis buffer 3	1 x 125 mL		1x125 mL	Ready to use
ADIAMAG beads	Magnetics beads	2 x 1.5 mL	1 x 12 mL		Ready to use
Binding Buffer - B2	Binding buffer	1 x 180 mL	1 x 500 mL		Ready to use
Wash Buffer - W3	Wash Buffer	1 x 75 mL	1 x 300 mL		Ready to use
Wash Buffer - W4	Wash Buffer	1 x 75 mL	1 x 300 mL		Ready to use
Elution Buffer - E6	Elution Buffer	1 x 30 mL	1 x 125 mL		Ready to use
Proteinase K - PK	Enzyme	1 x 75 mg	4 x 75 mg		Lyophilized – to rehydrate
Proteinase Buffer - BPK	Rehydration buffer	1 x 8 mL	1 x 15 mL		Ready to use

2. Validity and storage

After reception, all the extraction reagents can be stored at room temperature (+18 to +25 °C) and are stable for up to 1 year and at the most until the shelf-life of the kit. Leave the flasks closed to prevent evaporation.

Prior to use, resuspend the lyophilized proteinase K using 2.6 mL of “Proteinase Buffer - BPK”. The Proteinase K solution should be stored at < -15 °C.

In case the LB2 Lysis Buffer precipitate, preheat the solution at +70 °C until it clears.

In case the LB1 and LB3 Lysis Buffer precipitate, preheat the solution at +30-40 °C until it clears.

3. Equipment required, but not supplied in the kit

Warning: The material should be Nuclease-free (e.g., autoclaved 25 minutes twice at +121 °C or once 60 minutes at +121 °C)

- Class II Microbiological Safety Cabinet.
- Centrifuge for microtubes and 20 mL tubes.
- Grinder (Mixer Mill or Fast Prep).
- Incubator, heating bath or block heater.
- Vortex.
- 1 - 10 µL pipette, 20 - 200 µL pipette and 200 – 1 000 µL pipette.
- Nuclease-free filter tips.
- Nuclease-free microtubes: 1.5 mL and 2 mL.
- Sterile tubes of 10 or 15 mL.
- Powder-free latex or nitrile gloves.
- Razor blades.
- 80 % ethanol solution.
- Sterile distilled water.
- Sterile saline water (NaCl 8.5 g/L)
- 1X PBS buffer (recommended composition, NaCl 150 mM, Na₂HPO₄ 5 mM, KH₂PO₄ 1.7 mM, without Ca²⁺, without K⁺ - another composition can be used after validation by the user).
- MEM medium + antibiotics (penicillin 100 IU/mL and streptomycin 100 µg/mL).

Specific equipment for KingFisher instruments**KingFisher 96/Flex**

- DEEP WELL plates (Thermo scientific, 96 tests: ref. 10373480).
- ELUTION PLATES (Thermo scientific, 96 tests: ref. 10357939).
- TIPS (Thermo scientific, 96 tests: ref. 11744978).

KingFisher DUO

- DEEP WELL plates (Thermo scientific, 96 tests: ref. 10373480).
- TIPS-12 (Thermo scientific, 600 tests: ref. 97003500).

KingFisher mL

- KingFisher combi 240 (Thermo Scientific, 240 tests: ref :97002141).

Specific equipment for *Mycobacterium avium* subsp. *paratuberculosis* detection

- ADIAFILTER (Bio-X Diagnostics, 100 tests: ref. ADIFIL100).
- ADIAPREP (Bio-X Diagnostics, 100 tests: ref. ADPREP-200).
- Glass beads for grinding only for Mixer mill:
 - ADIAPURE™ ALIQUOTED GLASS BEADS (Bio-X Diagnostics, 480 tests: ref ADIADPBA1-480).
 - ADIAPURE™ GLASS BEADS RACKS 4x96 (Bio-X Diagnostics, 384 tests: ref ADPBIA-4x96).
- Glass beads for grinding only for grinder Fast Prep or similar:
 - Lysing Matrix B tubes (MP Biomedical, 100 extractions: ref. 116911.100).

Specific equipment for SBV virus or Influenza virus detection from tissue

- Lysing Matrix D tubes (MP Biomedicals, 100 extractions: ref. 116913.100) only for grinder Fast Prep.
- Metal beads 3 mm (e.g., Qiagen, 200 extractions: ref. 69997) only for grinder Mixer Mill.

III. Use of samples and controls

1. Precautions

Caution:

Prepare buffers according to the §II.2.

Buffers can contain toxic substances, please consult the MSDS safety data sheet.

Store reagents at the recommended temperature.

Only appropriately trained personnel should perform this extraction. Ensure the micropipettes used are calibrated. The quality of the obtained results depends on rigorous respect of good laboratory practices.

PCR generates large amount of DNA. A few molecules of amplified products are sufficient to generate a positive result. Hence, PCR tubes should not be opened after amplification.

Samples for analysis should be handled and disposed of as biological waste. **Take all measures of security and confinement required for the manipulation of the concerned biological agents.**

Before starting the process, read the entire protocol and scrupulously respect it.

2. Storage of nucleic acid extracts

Extracted DNAs can be stored at 4 °C for up to 24 hours after extraction. For long term storage, DNAs should be kept at < -15 °C.

Extracted RNAs are highly sensitive to temperature. Extraction should be performed at room temperature as fast as possible to avoid degradation. Crude extracts can be stored at 4 °C for a few hours after extraction. For long-term storage, RNAs should be stored at < -65 °C.

3. Controls preparation

Several controls should be included for each extraction.

The mix of the different controls included in the kits allows validation of all the steps (extraction and amplification) for all the samples.

- The endogenous or exogenous internal control included in the ADIAVET™ or ADIALYO™ kits allows validation of the extraction and amplification steps of each sample.
- The positive control included in the ADIAVET™ or ADIALYO™ kits allows validation of the specific target amplification.

Other controls should or must be added.

A. Negative control of extraction (required)

To verify the absence of cross-contamination, at least one negative control has to be included per trial (e.g., AFNOR NF U47-600-1 guidelines recommends 1 negative control per 24 columns or 4 negative controls per trial of 96-wells plate). This control can be a negative sample or a buffer used for dilution.

B. Positive control of extraction (recommended)

A positive control (including the specific pathogen) can be added in each trial. It can be a positive sample available in the laboratory or a negative sample spiked with the specific pathogen. This positive control will be detected close to the limit of detection of the method and allow comparison of the results obtained in different assays.

IV. Preparation of the buffers and loading of the instruments

1. Preparation of the binding buffer

We recommend preparing the binding buffer just before use.

Carefully mix the “ADIAMAG Beads” solution before each use.

During buffer preparation, always add one more reaction than needed (to cover the loss of solution during multi-pipetting).

Mix (per reaction):

- 600 µL of **Binding Buffer B2**.
- 13 µL of **ADIAMAG Beads**.

2. Loading of the instrument

According to the KingFisher™ instrument used, the loading is carried out as follow:

- KingFisher™ mL: one tube comb per sample.
- KingFisher™ DUO: one 96-plate.
- KingFisher™ 96/Flex: one 96-plate per step.

Distribute the buffer in each well/line/plate as mentioned in the following table:

	Buffer to add	KingFisher™ mL	KingFisher™ DUO	KingFisher™ 96/Flex
Preparation of reagents	600 µL binding buffer (600µL Binding Buffer B2 + 13 µL ADIAMAG beads)	Well 1 	Line B 	Plate 1 
	350 µL Wash Buffer W3	Well 2 	Line C 	Plate 2 
	350 µL Wash Buffer W4	Well 3 	Line D 	Plate 3 
	350 µL Ethanol 80 %	Well 4 	Line E 	Plate 4 
	60 µL or 100 µL* Elution Buffer E6	Well 5 	Line F 	Plate 5 
	Tip	On rail 	Line A 	Plate 6 
Loading of samples and run the instrument	Transfer the sample prepared according to §V, in the Well 1/Line B/Plate 1. Turn on the instrument. Select the program. Place the comb or plates in the instrument. Press Start. After extraction, collect the purified nucleic acids in Well 5/Line F/Plate 5.			

*100 µL for the influenza virus, PRV, *Besnoitia besnoiti*, Marek, Porcine Circovirus, *Mycobacterium avium* subsp *paratuberculosis* (with ADIAPREP protocol), *Salmonella* and fish diseases.

3. Instrument programs

All instrument programs listed below are compatible with the PCR kit range ADIAVET™ or ADIALYO™:

(Summarizing the standard and short programs for the different automates)

	Standard program (34 minutes)	Short program (21 minutes)
KingFisher™ 96/Flex	KF96V3-2	KF96V4-2
KingFisher™ DUO	KFDOUV3-1	KFDOUV4-1
KingFisher™ mL	KFMLV3	KFMLV4

The above programs, as well as programs for another instrument, are available on request (support.pcr@biox.com).

Warning, in the layout of our program, there may be a difference between real volume and programmed volume. This difference is validated.

V. Samples preparation before transfer on the device

1. From blood/sera

	Virus: BVDV, BTV, SBV, EHDV, PRRSV, CSFV, FIP and ASFV	PCV2 and PCV3	<i>A. phagocytophilum</i> and <i>Besnoitia besnoiti</i>
Sample preparation	Take 100 µL of sample.	Take 100 µL of sample.	Place 500 µL to 1 mL of bovine blood or 100 µL of equine blood in a microtube. Add 1 mL of sterile distilled water . Mix and incubate 10 minutes on ice. Centrifuge 6 000 g/5 minutes. Discard the supernatant. Add 1 mL of sterile distilled water . Centrifuge 6 000 g/5 minutes. Discard the supernatant.
Lysis	Add 250 µL of Lysis Buffer LB2 . ² Mix.	Add 250 µL of Lysis Buffer LB2 + 10 µL of Proteinase PK . ¹ Mix. Incubate 15 minutes at room temperature.	
Loading of the instrument	Transfer 200 µL in the binding buffer (well 1/line B/plate 1). ³		

¹A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL** of **EPC-Ext** provided in the kit of the pathogen of interest.

Refer to the corresponding package insert for preparation, storage and use of the control.

³ For 96-plate lysis, the binding buffer can be added directly to the lysed sample.

2. From oral fluids

	<i>M. hyopneumoniae</i> , PRRS virus and Influenza virus
Sample preparation	Take 100 µL of sample.
Lysis	Add 250 µL of Lysis Buffer LB2 . ¹ Mix.
Loading of the instrument	Transfer 200 µL in the binding buffer (well 1/line B/plate 1). ²

¹ Add **5 µL** of **EPC-Ext** provided in the kit of the pathogen of interest.

Refer to the corresponding package insert for preparation, storage and use of the control.

² For 96-plate lysis, the binding buffer can be added directly to the lysed sample.

3. From milk

	BVDV, <i>C. burnetii</i> and <i>Chlamydia</i>
Sample preparation	Take 100 µL of sample.
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK. ^{1,2} Mix. Incubate 15 minutes at 56 °C.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/line B/plate 1).

¹A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL of EPC-Ext** provided in the kit of the pathogen of interest.

Refer to the corresponding package insert for preparation, storage and use of the control.

4. From culture supernatant

	PRV, IHN, VHSV and IPNV
Sample preparation	Take 100 µL of sample.
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK. ^{1,2} Mix. Incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the whole volume in the binding buffer. (Well 1/line B/plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL of EPC-Ext** provided in the kit of the pathogen of interest.

Refer to the corresponding package insert for preparation, storage and use of the control.

5. From ear notch samples and humid TST-L from ALLFLEX ear notch (BAH)

	BVDV
Sample preparation	From 1 ear tissue sample , dry or humid. <i>In the case of an ALLFLEX TST-L loop, remove the preservation liquid from the collection tube.</i>
Lysis	Add 350 µL of Lysis Buffer LB3. ¹ Mix. Incubate 15 minutes at room temperature. ²
Loading of the instrument	Transfer 100 µL of individual or pooled³ supernatant in the binding buffer (well 1/Line B/plate 1).

¹ Add **5 µL of EPC-Ext** provided in the kit of the pathogen of interest.

Refer to the corresponding package insert for preparation, storage and use of the control.¹

² For new analysis, each individual supernatant can be stored at +4 °C for 24 hours or at < -15 °C for long-term

³ Up to 25 samples can be pooled together. Mix 50 µL of each sample and homogenize.

6. From conservation buffer of TST-L from ALLFLEX ear notch (BAL)

In the case of ALLFLEX TST-L, eject the biopsy into the preservation buffer and adjust, if necessary, the volume to 250 µL with ALLFLEX preservation buffer and incubate for 1 hour at room temperature. For extraction, use the buffer as matrix.

	BVDV
Sample preparation	Transfer 100 µL of conservation buffer or pool ¹ of conservation buffer from ear notch sample.
Lysis	Add 100 µL of LB1 Lysis Buffer + 10 µL of Proteinase PK. ² Mix. Incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ Up to 25 samples can be pooled together. Mix 50 µL of each sample and homogenize.

² A pre-mix can be prepared just before use and added to each sample.

7. From skin biopsy

	<i>Besnoitia besnoiti</i>
Sample preparation	Place 50 mg of skin biopsy in a microtube.
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK + 100 µL PBS Buffer 1X (or sterile saline water) ¹ Mix. Incubate a night at +56 °C.
Loading of the instrument	Transfer the whole volume in the binding buffer (Well 1/Lane B/Plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

8. From trachea- bronchial washing

	<i>M. hyopneumoniae</i> , PRRS virus and Influenza virus
Sample preparation	Transfer 1 mL of trachea-bronchial washing in microtube. Centrifuge 30 minutes at 10 000g. Discard the supernatant.
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK. ^{1,2} Mix and incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

² Add 5 µL of EPC-Ext provided in the kit of the pathogen of interest.

Refer to the corresponding package insert for preparation, storage and use of the control.

9. From water and urine

	<i>Leptospira</i>
Sample preparation	Transfer 10 mL in a tube. Centrifuge 30 minutes at 10 000 g or 10 minutes at 4 500 g. Discard the supernatant. Add 1 mL PBS 1X. Mix.
Lysis	Transfer 100 µL in a microtube. Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK. ¹ Mix and incubate 15 minutes at +56 °C.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

10. From coelomic liquid

	IPNV
Sample preparation	Take 100 µL of sample.
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK + 5 µL of EPC-Ext. ¹ Mix. Incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ Specific EPC-Ext are included in each kit.

See the kit of interest instruction manual to prepare, store and use this control.

A pre-mix can be prepared just before use and added to each sample.

11. From foetal gastric fluid

	<i>C. burnetii, Chlamydia, Salmonella</i>
Sample preparation	Take 100 µL of sample. (If the liquid is difficult to collect, dip a swab and treat it according to the swab protocol §15.A.f).
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK. ^{1,2} Mix and incubate 15 minutes at +56 °C.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL of EPC-Ext** included in the kit.

Refer to the corresponding package insert for the preparation, storage and use of the control.

12. From feather

	Avian Influenza virus, Marek virus
Sample preparation	Cut the calamus of 1 to 5 feathers precociously to avoid any projections and place them in 2 mL of physiological saline. Take 100 µL of sample.
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK . ^{1,2} Mix. Incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL** of **EPC-Ext** included in the kit.

Refer to the corresponding package insert for the preparation, storage and use of the control.

13. From FTA Card

	Avian Influenza virus, Avian mycoplasmas
Lysis	Cut a 3mm ² piece of the FTA Card. Place it in a microtube. Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK . ^{1,2} Mix and incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the whole volume in the binding buffer (well 1/Line B/plate 1).

¹ A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL** of **EPC-Ext** included in the kit.

Refer to the corresponding package insert for the preparation, storage and use of the control.

14. From tissue or brain

A. Sample preparation

a) BVDV, CSFV, ASFV viruses (lymphoid tissues: spleen, ganglions, tonsil) and PRV virus (brain, lung) with Lysis buffer LB3.

Place **20 mg** of sample in a microtube.

Continue according to the table below.

b) BTV, EHDV, SBV (brain, spleen), influenza virus and Marek virus

Grind^{1 or 2} **0.1 g** of sample with **1 mL** of **PBS Buffer 1X** or **sterile saline water**.

Transfer **100 µL** of sample in a microtube.

Continue according to the table below.

¹ For example with a grinder like Mixer Mill: add a metal bead (3 mm) and grind 2 minutes at 30 Hz then centrifuge 6 000 g/2 minutes.

² For example with a grinder like Fast Prep: in a Lysing Matrix D tube, grind twice 20 seconds at 6m/sec with a 5 minutes break on ice between both. Then centrifuge 2 000 g/3 minutes.

c) IHN, VHSV and IPNV (spleen, anterior kidney and heart or brain)

Homogenize, by stomacher, mixer or mortar and pestle with sterile sand, re-suspend in the original transport medium with a ratio of 10 % w / v, centrifuge for 15 minutes at 4 000 g.

or

Grind^{1 or 2} **0.1 g** of sample with **1 mL** of **PBS Buffer 1X** or **sterile saline water**.

Transfer **100 µL** of sample in a microtube.

Continue according to the table below.

¹ For example with a grinder like Mixer Mill: add a metal bead (3 mm) and grind 2 minutes at 30 Hz then centrifuge 6 000 g/2 minutes.

² For example with a grinder like Fast Prep: in a Lysing Matrix D tube, grind twice 20 seconds at 6m/sec with a 5 minutes break on ice between both. Then centrifuge 2 000 g/3 minutes.

d) *M. hyopneumoniae*, PRRSV (lung), ASFV (lymphoid tissues: spleen, ganglions, tonsil) viruses, PCV2 and PCV3

Place **20 mg** of sample in a microtube.

NB: for the PRRS virus, pooled analysis (up to 3 samples) can be performed, n x 20 mg.

Add **1 mL** of **sterile saline water**.

Grind (e.g., with a grinder like Mixer Mill, add a metal bead (3 mm) and grind 2 minutes at 30 Hz).

Centrifuge 6 000 g/2 minutes.

Transfer **100 µL** of the supernatant of the grinded sample in a microtube.

Continue according to the table below.

e) *Actinobacillus pleuropneumoniae*

Vortex one biopsy in **1 mL** of **sterile saline water**.

Place **20 µL** of obtained liquid in a microtube.

Continue according to the table below.

f) *A. phagocytophilum*, *C. burnetii*, Chlamydia, Leptospira., *N. caninum*, *T. gondii* and Salmonella (tissue e.g., cotyledon of placenta, foetal tissues)

The analysis from spleen is not recommended. Potential PCR inhibitors can disrupt the analysis.

Rub within the tissue using dry swab.

Continue according to § 15.A.f.

g) *N. caninum* and *T. gondii* (brain)

Mix **one volume of brain** and **one volume of sterile saline water**.

Vortex.

Transfer **100 µL** of the obtained liquid in a microtube.

Continue according to the table below.

h) *M. avium* subsp. *paratuberculosis*

Place **20 mg** of **sample** in a microtube.

Add **1 mL** of **sterile saline water**.

Grind 10 minutes at 30 Hz (e.g., with a grinder like Mixer Mill) with a metal bead (3 mm) and 300 mg of glass beads (Bio-X Diagnostics, 480 tests: ref ADIADPBA1-480).

Centrifuge 3 000 g/5 minutes.

Transfer **100 µL** of **the supernatant of the grinded sample** in a microtube.

Continue according to the table below.

B. Nucleic acids extraction and purification

	BVDV Virus	PRV, CSFV and ASFV viruses	<i>M. hyopneumoniae</i> , SBV, BTV, EHDV PRRSV, influenza, Marek, IHN, VHSV, IPNV, ASFV viruses, PCV2 and PCV3	<i>A. pleuropneumoniae</i> , <i>N. caninum</i> and <i>T. gondii</i>	<i>M. avium</i> subsp. <i>paratuberculosis</i>
Lysis	Add 350 µL of Lysis Buffer LB3 . ⁴ Mix.		Add 100 µL of Lysis Buffer LB1 + 10 µL of proteinase PK . ^{1,4} Mix.		
	Incubate 15 minutes at room temperature.	Grind. ²⁻³	Incubate 15 minutes at room temperature.	Incubate 15 minutes at +56 °C.	No incubation.
Loading of the instrument	Transfer 100 µL in the binding buffer (well 1/Line B/plate 1).		Transfer the totality in the binding buffer (well 1/Line B/plate 1).		

¹A pre-mix can be prepared just before use and added to each sample.

²For example with a grinder like Mixer Mill: add a metal bead (3 mm) and grind 2 minutes at 30 Hz then centrifuge 6 000 g/2 minutes.

³For example with a grinder like Fast Prep: in a Lysing Matrix D tube, grind twice 20 seconds at 6m/sec with a 5 minutes break on ice between both. Then centrifuge 2 000 g/3 minutes.

⁴Add **5 µL** of **EPC-Ext** included in the kit.

Refer to the corresponding package insert for the preparation, storage and use of the control

15. From swab

A. Sample preparation

a) *M. hyopneumoniae*, PRRSV virus and PRV virus.

Add **2 mL** of **sterile saline water** in the tube with the swab.

Vortex.

NB: for PRRSV virus, up to 3 swabs can be pooled together. Transfer the supernatant obtained in the tube with the next swab.

Press each swab to collect as much liquid as possible.

Transfer the liquid in a 2 mL-microtube.

Place **100 µL** of **the obtained liquid** in a microtube.

Continue following the table below.

b) FIP (feline coronavirus).

Add **2 mL** of **sterile saline water** in the tube with the swab.

Vortex.

Press each swab to collect as much liquid as possible.

Transfer the liquid in a 2 mL-microtube.

Place **100 µL** of **the obtained liquid** in a microtube.

Continue according to the table below.

c) Avian Influenza Virus

Add 1 swab in a tube with **1 mL** of **MEM medium** (+antibiotic if viral culture is needed) or **sterile saline water**.

Vortex.

Up to 10 swabs can be pooled together (mix 1 volume of each individual). Keep the individual mixtures at a temperature < -65 °C.

In case of pools, some weak positive samples can be not detected

Take **100 µL** of **the obtained liquid**.

Or

Add **1 to 5 swabs** into a tube with **2 mL** of **MEM medium** (+antibiotic if viral culture is needed) or **sterile saline water**.

Vortex.

Up to 10 swabs can be pooled together (mix 1 volume of each individual).

Take **100 µL** of **the obtained liquid**.

Continue according to the table below.

d) Swine Influenza Virus

Add 1 swab into a tube with **2 mL** of **MEM medium + antibiotic** (to allow subsequent viral culture) or **sterile saline water**.

Vortex.

Take **100 µL** of **the obtained liquid**.

Continue according to the table below.

e) *Actinobacillus pleuropneumoniae*

Vortex one swab in **1 mL** of **sterile saline water**.

Place **20 µL** of **the obtained liquid** in a microtube.

Continue according to the table below.

f) *A. phagocytophilum*, *C. burnetii*, Chlamydia, Leptospira., *N. caninum*, *T. gondii*, BoHV-4, ASFV and Salmonella

Vortex one swab in **1 mL** of **PBS Buffer 1X**.

Place **100 µL** of **the obtained liquid** in a microtube.

Continue according to the table below.

g) Avian mycoplasmas, Ornithobacterium rhinotracheale and Infectious Bronchitis virus

Cut 1 to 3 swabs in **1 mL** of **sterile saline water**.

Place **100 µL** of the obtained liquid in a microtube.

Continue according to the table below.

h) CEM: Taylorella equigenitalis, Taylorella asinigenitalis, Klebsiella pneumoniae and Pseudomonas aeruginosa

Vortex one swab in **500 µL** of **PBS Buffer 1X**.

Place **100 µL** of **the obtained liquid** in a microtube.

Continue according to the table below.

B. Nucleic acids extraction and purification

	<i>M. hyopneumoniae</i> , FIP, PRRSV, PRV, influenza viruses and Infectious Bronchitis virus	<i>A. pleuropneumoniae</i> , <i>A. phagocytophilum</i> , <i>C. burnetii</i> , <i>Chlamydia</i> , <i>Leptospira.</i> , <i>N. caninum</i> , <i>T. gondii</i> , <i>Salmonella</i> , BoHV-4, Avian mycoplasma, <i>Ornithobacterium rhinotracheale</i> , CEMO and ASFV
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of proteinase PK . ^{1,2}	
	Mix.	
	Incubate 15 minutes at room temperature.	Incubate 15 minutes at +56 °C.
Loading of the instrument	Transfer the totality in the binding buffer (well 1/Line B/plate 1).	

¹ A pre-mix can be prepared just before use and added to each sample.

² Add **5 µL** of **EPC-Ext** included in the kit of interest.

See the kit of interest user manual to prepare, store and use this control.

16. From faeces

A. Sample preparation

a) *M. avium* subsp. *paratuberculosis*

Note: Overnight rehydration at room temperature may be recommended for **goat and sheep faeces**. **Environmental samples** (for example soil scrapings from different breeding areas, ...) have to be treated as faecal samples. Dilute 3-10 g of sample in water.

Three ADIAMAG protocols are applicable:

- With ADIAPREP (ref. ADPREP-200)

Collect 1 spoon of faecal matter and transfer it into the ADIAPREP.

Vortex until a homogeneous suspension is obtained.

Transfer 1 mL into a clean tube, centrifuge 5 minutes at 3000 g and discard the supernatant.

Add 300 mg grinding beads and 500 µL sterile deionized water to the pellet.

Grind for 5 minutes at 30 Hz on the Mixer Mill or 3 x 45 seconds on the Fast Prep/Ribolyser and centrifuge for 5 minutes at 3000 g.

Transfer 100 µL of sample in a microtube.

Continue according to the table below.

- With ADIAFILTER (ref. ADIFIL100)

Collect 3 g (+/- 0.2) of faecal matter, vortex into 20 mL of sterile deionized water (or identical weight/volume ratio) until sedimentation for 10 to 20 minutes.

Place 10 mL of supernatant in the ADIAFILTER, centrifuge for 5 minutes at 3000 g and discard the supernatant and ADIAFILTER.

Add 500 µL sterile deionized water to the pellet, vortex and transfer to a microtube containing 300 mg grinding beads.

Grind for 10 minutes at 30 Hz on Mixer Mill or 3 x 45 seconds on Fast Prep/Ribolyser and centrifuge for 5 minutes at 15 000 g.

Transfer 100 µL of sample in a microtube.

Continue according to the table below.

- Without prefiltration system

Collect 3 g (+/- 0.2) of faecal matter (or identical weight/volume ratio), vortex into 20 mL of sterile deionized water until sedimentation for 10 to 20 min.

Transfer 1 mL to a microtube containing 300 mg of grinding beads.

Grind for 10 minutes at 30 Hz on Mixer Mill or 3 x 45 seconds on Fast Prep/Ribolyser and centrifuge for 5 minutes at 15 000 g.

Transfer 100 µL of sample in a microtube.

Continue according to the table below.

b) *C. burnetii*

Add **5 mL of PBS Buffer 1X** to **1 g of faeces**.

Homogenize, e.g., with a vortex, for at least 15 seconds.

Centrifuge 3 000 g at 2 minutes.

Transfer **100 µL of sample** in a microtube.

Continue according to the table below.

c) *T. gondii*

Weigh **1 g** of faeces in a previously labelled 10 mL or 15 mL sterile tube.

Add **10 mL of PBS Buffer 1X** (*This preparation is stable for 24 hours at room temperature*).

Vortex until a homogenous solution is obtained.

Allow to settle 2 to 5 minutes.

Transfer **500 µL of supernatant** in a previously identified microtube.

Centrifuge at 3 000 g for 5 minutes. Discard the supernatant.

Homogenise the pellet with **1 mL of PBS Buffer 1X** (*This solution is stable for 24 hours at room temperature*).

Transfer **500 µL of supernatant** in a tube containing 300 mg of glass beads.

Disrupt 10 minutes at 30 Hz with a Mixer Mill (or transfer the obtained solution in a Lysis Matrix B tube and disrupt 3 x 45 seconds at 4 m/sec with the Fast Prep).

Centrifuge at 15 000 g 5 minutes.

Transfer **100 µL of supernatant** in a microtube.

Continue according to the table below.

d) Influenza virus

Weigh **1 g** of manure and add **5 mL of physiological water**

Vortex until a homogenous solution is obtained.

Sediment 5 minutes.

Transfer **100 µL of supernatant** in a microtube.

Continue according to the table below.

B. Nucleic acids extraction and purification

	<i>M. avium</i> subsp. <i>paratuberculosis</i> (with ADIAPREP) ²	<i>M. avium</i> subsp. <i>paratuberculosis</i> (with ADIAFILTER) <i>C. burnetii</i> , <i>T. gondii</i>	<i>M. avium</i> subsp. <i>paratuberculosis</i> (without prefiltration system) Influenza virus
Lyse	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK + 5 µL of EPC-Ext. ¹ .		
	Mix.		
	No incubation.	Incubate 15 minutes at +56 °C .	Incubate 15 minutes at room temperature .
Chargement de l'automate	Transfer the totality in the binding buffer (well 1/Line B/plate 1).		

¹ Specific EPC-Ext are included in each kit. See the kit of interest instruction manual to prepare, store and use this control.

A pre-mix can be prepared just before use and added to each sample.

² Use 100 µL of E6 Buffer (well 5/Line F/plate 5).

17. From environmental sample (drag swab)

A. Sample preparation

a) *M. avium* subsp. *paratuberculosis*

Environmental samples (dung scrapings, holding area...) are treated as faeces. Continue to § 16.A.a).

b) Influenza Virus, Marek virus and Infectious Bronchitis virus

Mix drag swab in a maximum volume of 40 mL of STP or PBS 1X Buffer. 70 mL for an overshoe.

Centrifuge 1 mL of supernatant 10 to 15 minutes at 800 à 1 200g.

Transfer **100 µL** of **supernatant** in a microtube.

Continue according to the table below.

B. Nucleic acids extraction and purification

	Marek virus	Influenza virus Infectious Bronchitis virus
Lysis	Add 100 µL of Lysis Buffer LB1 + 10 µL of Proteinase PK + 5 µL of EPC-Ext. ¹ Mix.	
	incubate 15 minutes at +56 °C.	incubate 15 minutes at room temperature.
Loading of the instrument	Transfer the totality in the binding buffer (well 1/Line B/plate 1).	

¹ Specific EPC-Ext are included in each kit. *See the kit of interest instruction manual to prepare, store and use this control.*

A pre-mix can be prepared just before use and added to each sample.

18. From sperm or egg fish

A. Sample preparation

a) IHNV, VHSV, IPNV (liquid sperm)

Mix **one volume of sperm** and **one volume of PBS 1X**.

Vortex 20 seconds.

Centrifuge 2 minutes at 6 000 g.

Transfer **100 µL of the obtained liquid** in a microtube.

Continue according to the table below.

b) IHNV, VHSV, IPNV (coagulated sperm, egg)

Homogenize, by stomacher, mixer or mortar and pestle with sterile sand, re-suspend in the original transport medium with a ratio of 10 % w / v, centrifuge for 15 minutes at 4 000 g.

or

Grind ^{1 or 2} **0.1 g of sample** with **1 mL of PBS Buffer 1X** or **sterile saline water**.

Transfer **100 µL of sample** in a microtube.

Continue according to the table below.

¹ For example with a grinder like Mixer Mill: add a metal bead (3 mm) and grind 2 minutes at 30 Hz then centrifuge 6 000 g/2 minutes.

² For example with a grinder like Fast Prep: in a Lysing Matrix D tube, grind twice 20 seconds at 6m/sec with a 5 minutes break on ice between both. Then centrifuge 2 000 g/3 minutes.

B. Nucleic acids extraction and purification

	IHNV, VHSV, IPNV
Lysis	Add 100 µL of Lysis Buffer LB1 +10 µL of Proteinase PK + 5 µL of EPC-Ext. ¹ Mix. Incubate 15 minutes at room temperature .
Loading of the instrument	Transfer the totality in the binding buffer (well 1/Line B/plate 1).

¹ Specific EPC-Ext are included in each kit. See the kit of interest instruction manual to prepare, store and use this control.

A pre-mix can be prepared just before use and added to each sample.

VI. Amplification

For the amplification of extracted nucleic acids, please refer to “Amplification” and “Interpretation of results” paragraphs of the ADIAVET™ or ADIALYO™ kit of interest user manual.

VII. Index of symbols

Symbol	Meaning
	Catalog number
	Manufacturer
	Upper temperature limit
	Use by date
	Batch code
	Consult Instructions for Use
	Contains sufficient for <n> tests
	Keep away from sunlight
	For veterinary in vitro use only

Bio-X Diagnostics, the logos, ADIAGENE, ADIAPURE™ and ADIAVET™ are used, pending and/or registered trademarks belonging to ADIAGENE and/or Bio-X Diagnostics, or one of its subsidiaries, or one of its companies. Any other name or trademark is the property of its respective owner.



**ADIAGENE S.A.S.**
9, rue Gabriel Calloët-Kerbrat
22440 Ploufragan - France

RCS 417 876 299
Tel. 33 (0)2 96 68 40 20
www.biox.com