



Adia^X Control

Extraction Positive Control SALMO

Reference: ADC12EPC

Salmonella enterica serovar *Typhimurium* and *Salmonella enterica* serovar *Enteritidis* positive control for extraction control

For veterinary *in vitro* use only



Kit composition

| Content | | ADC12EPC Kit 100 reactions |
|-----------|---|---|
| EPC SALMO | Extraction Positive Control <i>Salmonella</i> | 1 lyophilized vial (To reconstitute) |
| NF-Water | Nuclease free water | 1 x 1 mL tube (Ready to use) |

Associated PCR kit(s)

| Associated kit(s) | Reference(s) |
|--------------------------------|--------------|
| ADIAVET™ SALMO FAST TIME | ADI122-100 |
| ADIAVET™ SALMO ST-SE FAST TIME | ADI631-100 |

Revision history

| Date | Version | Modifications |
|---------|---------|--|
| 06/2024 | V01 | First version |
| 09/2024 | V02 | Modification of the NF-Water tube from 500 µL to 1 mL. |

Note: minor typographical, grammar and formatting changes are not included in the revision history.

A. Test principle

The extraction positive control is made with a pool of inactivated, lyophilized culture of *Salmonella enterica typhimurium* and *Salmonella enterica enteritidis*.

This extraction positive control after rehydration can be used for two purposes:

- the extraction positive control can be used within each extraction series and allows the day-to-day extraction step control.
- it can be used to confirm the LOD_{METHOD}. It is calibrated between 1 and 100 the LOD_{METHOD}.

It is the responsibility of the laboratory to obtain the matrix free of *Salmonella*.

B. Storage conditions

On receipt, the kit should be stored dry at +2/8 °C.

Reconstituted reagents should be stored at <-15 °C until the expiration date.

Store away from sunlight.

Do not mix reagents of two different batches.

C. Material required but not provided

- Real-time Thermal cycler and device.
- Instrument for homogenous mixing of tubes.
- Pipettes of 1 - 10 µL, 20 - 200 µL and 200 - 1000 µL.
- Nuclease-free filtered pipette tips.
- Nuclease-free microtubes of 1,5 mL and 2 mL.
- Powdered-free latex or nitrile gloves.
- Nuclease-free water.
- Associated PCR kits.
- Kit for nucleic acids extraction.
- Matrix free of *Salmonella*.

D. Warnings and precautions

- For veterinary in vitro use only.
- For animal use only.
- For professional use only.
- All instructions must be read before performing the test and strictly respected.
- Do not use reagents after the expiration date.
- Do not use reagents if the packaging is damaged.
- Do not open PCR wells or tubes after amplification.
- Do not mix reagents from different batches.
- Used material must be disposed of in compliance with the legislation in force regarding environmental protection and biological waste management.
- This kit contains products of animal origin. Certified knowledge of the origin and/or sanitary state of the animals does not totally guarantee the absence of transmissible pathogenic agents. It is therefore recommended that these products be treated as potentially infectious and handled observing the usual safety precautions (do not ingest or inhale).

E. Procedure

1. Reconstitution of the control

- Add **500 µL** of « **NF Water** » in « **EPC SALMO** » vial.
- Vortex vigorously at least 20 seconds.
- After reconstitution, the control is calibrated between **1 et 100X** LD_{METHOD}.
- Aliquot and store at <-15°C. Do not defrost more than 3 times.

2. LOD_{METHOD} confirmation

a. Preparation

Dilute extemporaneously to 1/10 the « **EPC SALMO** » with the NF Water. *Dilution should not be stored.*

b. Extraction

Add **5 µL** of « **EPC SALMO x1/10** » to the matrix free of *Salmonella enterica spp* and extract **on duplicate** according to the used protocol. Two independent sessions should be realized.

c. Amplification

Nucleic acids are extracted with the associated Bio-X Diagnostics PCR kits according to kit instructions.

d. Validation of results

The four results obtained with « **EPC SALMO** » must be positive.

3. Extraction step control

a. Extraction

Add **5 µL** of « **EPC SALMO** » to the matrix free of *Salmonella enterica spp* and extract according to the used protocol.

b. Amplification

Nucleic acids are extracted with the associated Bio-X Diagnostics PCR kits according to kit instructions.

c. Validation of results

All the obtained results during the different series of extraction constitute the control chart.

Bibliography

- U47-600-1: Méthodes d'analyse en santé animale - PCR (réaction de polymérisation en chaîne) - Partie 1: exigences et recommandations pour la mise en œuvre de la PCR en santé animale.

Symbols

| Symbol | Signification |
|---|--|
|  | Catalog number |
|  | Manufacturer |
|  | Temperature limitation |
|  | Use by |
|  | Batch code |
|  | Consult Instructions for Use |
|  | Contain sufficient for "n" tests |
|  | For veterinary in vitro use only – For animal use only |
|  | Keep away from sunlight |